



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

AMRESKO LLC

28600 Fountain Parkway
Solon, Ohio USA 44139
440/349-1199 FAX: 440/349-1182
www.amresco-inc.com
Email: info@amresco-inc.com

CERTIFICATE OF QUALITY / CERTIFICATE OF ANALYSIS

Trypan Blue Solution, 0.4%

Code: K940

Chemical Formula:	Not Applicable	Manufacture Date:	(batch specific)
Molecular Weight:		Expiration/Reassay Date:	(batch specific)
CAS #:	Not Applicable		
Appearance:		Storage:	Cold
Dark blue liquid.		Grade:	TISSUE CULTURE GRADE

Additional Information

TEST	SPECIFICATION	DISPOSITION
Abs.@ Lambda Max (1:100, Water)	1.0000 - 1.3000	PASS
Lambda Max (1:100, Water)	585 - 595 nm	PASS

Spec Set: K940TCG

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Internal ID #: 269

Signature: _____

Title:

Date Printed:

08/10/2016

Page 1 of 1